

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	1865			
	CITRIC ACI	D 1:1 50% W/W SOLUTIC	N	
Test		Target/UOM	Range	Result
PREPARED TO FORMULA	TION ON	YES		YES
FILE		YES		
Sample ID				
Tested 02	2/22/24			
Tester JF	RIO			
Recorded 02	2/22/24			
CALCULATED W/W% CITR		50.0	49.8 - 50.2	50.0
		W/W%		
Sample ID				
Tested 02	2/22/24			
Tester JF	RIO			
Recorded 02	2/22/24			
APPEARANCE AND COLO	R CLI	EAR TO PALE YELLOW		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
-	2/22/24			
	RIO			
	2/22/24			
INSTRUMENTS USED DUR	RING	INSTRUMENT		M25,T246
PREPARATION				
Sample ID				
	2/22/24			
	RIO			
	2/22/24			
EXPIRATION DATE		MM/DD/YY		08/30/25
		MM/DD/YYYY		
Sample ID				
	2/22/24			
	RIO			
	2/22/24			
SHELF LIFE OF PRODUCT		SHELF LIFE		18 MONTHS
		TIME		-
Sample ID				
	2/22/24			
	RIO			
	2/22/24			
Lot # 4022125	5			
_01.0				
Made 02/22/2	14			



Page: 2 Date: 02/22/24 at 8:27 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager