

**Made** 02/19/24

Page: 1

Date: 02/19/24 at 2:22 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: HAC050F			
		PPM TOTAL NITROGEN 20	0 PPM SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	02/19/24			
Tester	AROS			
Recorded	02/19/24			
PPM CARBON		500	485 - 515	503
		PPM		
Sample ID				
Tested	02/19/24			
Tester	AROS			
Recorded	02/19/24			
CALCULATED PPM NIT		200.0	198.0 - 202.0	200.0
OALOOLA I LO I I III III	moden	PPM	100.0 202.0	200.0
Sample ID		1 1 101		
Tested	02/19/24			
Tester	02/19/24 AROS			
	02/19/24			
Recorded  APPEARANCE AND CO		CLEAR LIQUID		PASS
APPEARANCE AND CO	DLOR			PASS
0		APPEARANCE		
Sample ID				
Tested	02/19/24			
Tester	AROS			
Recorded	02/19/24			
INSTRUMENTS USED	DURING	INSTRUMENT		M108,VF2-B12
PREPARATION				
Sample ID				
Tested	02/19/24			
Tester	AROS			
Recorded	02/19/24			
EXPIRATION DATE		MM/DD/YYYY		05/30/2024
		MM/DD/YYYY		
Sample ID				
Tested	02/19/24			
Tester	AROS			
Recorded	02/19/24			
Lot # 4021				
LOI# 4021				



Page: 2

Date: 02/19/24 at 2:22 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**