

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4150 HYDROCHLORIC ACID 0.2 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
Actual Normality	0.2000 EQ/L	0.1995 - 0.2005	0.1996
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
N.I.S.T. Traceable to SRM 723	Yes YES		YES
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		02/20/2024
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	21.7
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		02/20/2024
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

AMBIENT TEST TEMPERATURE	20.0	15.0 - 30.0	21.7
DEG C	DEG C		
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
	APPEARANCE		
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
Actual Density	1.00187	1.00179 - 1.00194	1.00194
	GM/ML		
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T-164,M55,BID02259,PID M91
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
EXPIRATION DATE	MM/DD/YYYY		02/28/2026
	MM/DD/YYYY		
Sample ID			
Tested 02/20/24			
Tester JENC			
Recorded 02/20/24			
Lot #	4021659		
Made	02/20/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager