**Made** 02/27/24

Page: 1

Date: 02/27/24 at 9:59 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Customer PO#.				
Packag	ed Product: FIS	111F		
		fate Standard 10.0 ppm $w/v$	Solution	
Test		Target/UOM	Range	Result
	ED TO FORMULATION C	N YES		YES
FILE		YES		
	Sample ID			
	Tested 02/27/24			
	Tester AROS			
	Recorded 02/27/24			
CALCUL	ATED PPM SULFATE (Se	<b>04)</b> 10.00	9.95 - 10.05	10.00
		PPM		
	Sample ID			
	Tested 02/27/24			
	Tester AROS			
	Recorded 02/27/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 02/27/24			
	Tester AROS			
	Recorded 02/27/24			
INSTRUM	MENTS USED DURING	INSTRUMENT		P78,VF500-X42
PREPAR	ATION			
	Sample ID			
	Tested 02/27/24			
	Tester AROS			
	Recorded 02/27/24			
EXPIRATION DATE		MM/DD/YYYY		08/30/2024
		MM/DD/YYYY		
	Sample ID			
	Tested 02/27/24	ı		
	Tester AROS			
	Recorded 02/27/24	ı		
	ot # 4022769			

**Quality Manager**