

Page: 1

Date: 02/26/24 at 2:56 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	TEN222F			
		INDEX STANDARD 18 - 20	MG/100G (1 LITER FORMUL	ATION)
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES	-	YES
FILE		YES		
Sample ID				
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
BROMINE INDEX VALUE		19.00	15.00 - 24.00	19.72
		BROMINE IN		
Sample ID				
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
APPEARANCE AND CO		CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
INSTRUMENTS USED D		LAB EQUIPMENT USED		M100,M108,
PREPARATION		INSTRUMENT		,,
Sample ID				
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
EXPIRATION DATE	02/20/21	MM/DD/YY		05/30/24
EXI IIIAIION BAIL		MM/DD/YYYY		33/33/21
Sample ID		, 55, 1111		
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
SHELF LIFE OF PRODU		3 MONTHS		3 MONTHS
5E. E. E OI I HODO		TIME		5O.VIII 0
Sample ID				
Tested	02/26/24			
Tester	MMAT			
Recorded	02/26/24			
Lot # 4022638				
Made 02/26/24				
linuae -=/2	•			



Page: 2

Date: 02/26/24 at 2:56 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager