

Page: 1

Date: 02/28/24 at 4:00 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 3860			
	FLUORID	E STANDARD 10 PPM (1ML	=0.01MG)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	02/28/24			
Tester	JENC			
Recorded	02/28/24			
PPM FLUORIDE		10.0	9.8 - 10.2	10.0
		PPM		
Sample ID				
Tested	02/28/24			
Tester	JENC			
Recorded	02/28/24			
QCSP45:MEASURING I	FLUORIDE IN	YES		YES
WATER		QCSP45		
Sample ID				
Tested	02/28/24			
Tester	JENC			
Recorded	02/28/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		-		
Tested	02/28/24			
Tester	JENC			
Recorded	02/28/24			
INSTRUMENTS USED I		INSTRUMENT		P63,M79
PREPARATION	Johnid	III OTTIONILITI		1 00,1111 0
Sample ID				
Tested	02/28/24			
Tested	02/28/24 JENC			
Recorded	02/28/24			
EXPIRATION DATE	02/20/24	MM/DD/YYYY		09/30/2024
EXPINATION DATE		MM/DD/YYYY		03/30/2024
Sample ID				
Tested	02/28/24			
Tester	JENC			
Recorded	02/28/24			
Lot # 4022877				
Made 02/28/24				
wade 02/20), <u>4</u> 4			



Page: 2

Date: 02/28/24 at 4:00 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager