

Page: 1

Date: 03/01/24 at 10:35 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	3854			
	Fluorid	e Standard 50.0 ppm Ce	rtified Solution	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	03/01/24			
Tester	AFRE			
Recorded	03/01/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	03/01/24			
Tester	AFRE			
Recorded	03/01/24			
INSTRUMENTS USED DU	JRING	INSTRUMENT		P152,VF1-Z40,M116
PREPARATION				
Sample ID				
	03/01/24			
	AFRE			
	03/01/24			
PPM FLUORIDE		50.0	49.0 - 51.0	50.8
		PPM		
Sample ID				
	03/01/24			
	AFRE			
	03/01/24			
QCSP45:MEASURING FLUORIDE IN		YES		YES
WATER		QCSP45		
Sample ID				
-	03/01/24			
	AFRE			
	03/01/24			
EXPIRATION DATE	55/51/24	MM/DD/YY		09/30/24
		MM/DD/YYYY		33/33/21
Sample ID		, <i>55</i> , 1111		
	03/01/24			
	03/01/24 AFRE			
Recorded				
SHELF LIFE OF PRODUC		SHELF LIFE		6 MONTHS
SHEEF EN E OF FRODUC		TIME		OMORTHO
Sample ID		1 11VIL		
	03/01/24			
	03/01/24 AFRE			
	03/01/24			
Lot # 4022914				
Made 03/01/	724			



Page: 2

Date: 03/01/24 at 10:35 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager