

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6950			
Ŭ	CONDUCI	TIVITY STANDARD 1,000 uMH	IOS	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
CONDUCTIVITY @25 C.		1000	999 - 1001	1000
		US/CM		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
TRACEABLE TO N.I.S.T.	SRM 3191	YES		YES
		YES		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
APPEARANCE AND COLO	OR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
INSTRUMENTS USED DU	IRING	INSTRUMENT		T229,M92
PREPARATION				
Sample ID				
Tested	03/07/24			
	RSHE			
Recorded	03/07/24			
EXPIRATION DATE		MM/DD/YY		03/30/25
		MM/DD/YYYY		
Sample ID				
Tested	03/07/24			
Tester	RSHE			
Recorded	03/07/24			
SHELF LIFE OF PRODUC	т	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
	03/07/24			
	RSHE			
Recorded	03/07/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4030768
Made	03/07/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard St

Quality Manager