

Page: 1

Date: 03/07/24 at 6:00 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	DC989F	,		
	TOC SI	CANDARD 15.0 PPM CERTIFED	(PROPYLENE GLYCOL)	
Test		Target/UOM	Range	Result
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID	00/00/04			
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24	INICTEL INTENT		D00 D007 \/F0 \//7
INSTRUMENTS USED DURING		INSTRUMENT		P68,P397,VF6-W7
PREPARATION				
Sample ID				
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24			
EXPIRATION DATE		MM/DD/YY		05/11/24
		MM/DD/YYYY		
Sample ID				
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24			
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS
		TIME		
Sample ID				
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24			
Actual ppm T.O.C.		15.0	14.0 - 16.0	15.2
		PPM		
Sample ID				
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24			
Actual ppm T.O.C.		15.0	14.0 - 16.0	15.2
		PPM		
Sample ID				
Tested	03/06/24			
Tester	DTRA			
Recorded	03/06/24			
Lot # 40305	503			
Made 03/06				
WIAUE 00700	·, - 1			



Page: 2

Date: 03/07/24 at 6:00 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager