03/05/24

Made

Page: 1

Date: 03/05/24 at 1:08 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		D 0770 4E		nier PO#.	
Packag	ou i iouuoti	DC734F	######################################		
		AMMONIA S	TANDARD 30 PPM AS NH3		
Test			Target/UOM	Range	Result
EXPIRATION DATE		MM/DD/YYYY		05/10/2024	
			MM/DD/YYYY		
	Sample ID				
		/05/24			
		ROS			
		/05/24			
PREPARED TO FORMULATION ON		YES		YES	
FILE			YES		
	Sample ID				
		/05/24			
		ROS			
		/05/24			
CERTIFIE	ED PPM NH3		30.0	28.0 - 32.0	30.7
			PPM		
	Sample ID				
		/05/24			
		ROS			
		/05/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS	
			APPEARANCE		
	Sample ID				
		/05/24			
		ROS			
		/05/24			
INSTRUMENTS USED DURING		INSTRUMENT		P151,P121,VF3-J15,M107	
PREPAR	ATION				
	Sample ID				
	Tested 03	/05/24			
	Tester AF	ROS			
	Recorded 03	/05/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

lauland I