03/07/24

Made

Page: 1

Date: 03/07/24 at 2:23 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		Fluoride Standard 2.5 mg/L for EZ3000 Fluoride Analyzer			
Test		Targe	et/UOM	Range	Result
PREPARI	ED TO FORMULA		YES	•	YES
FILE		•	YES		
	Sample ID				
	Tested 0	3/07/24			
	Tester J	ENC			
	Recorded 0	3/07/24			
CALCULATED PPM FLUORIDE		RIDE	2.50	2.10 - 2.90	2.64
		PPM	I (MG/L)		
	Sample ID				
	Tested 0	3/07/24			
	Tester J	ENC			
	Recorded 0	3/07/24			
APPEARANCE AND COLOR		R CLEA	R LIQUID		PASS
		APPE	ARANCE		
	Sample ID				
	Tested 0	3/07/24			
	Tester J	ENC			
	Recorded 0	3/07/24			
INSTRUMENTS USED DURING		ring insti	RUMENT		P172,P118,VF5-E4
PREPAR	ATION				
	Sample ID				
	Tested 0	3/07/24			
	Tester J	ENC			
	Recorded 0	3/07/24			
EXPIRATION DATE		MM/E	DD/YYYY		06/30/2024
		MM/E	DD/YYYY		
	Sample ID				
	•	3/07/24			
	Tester J	ENC			
	Recorded 0	3/07/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I