



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1520			
BUFFER REFERENCE pH 9.0			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
PH OF PRODUCT/SOLUTION	9.000	8.980 - 9.020	9.009
	PH		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
QCSP 20 PH CALIBRATION METHODS	YES		YES
	METHOD		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
Actual Temp	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
N.I.S.T. Traceable	YES		YES
	YES		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
NIST TRACEABLE TO SRM186I AND 186II	YES		YES
	YES		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
PH NIST TRACEABLE TO SRM 187	YES		YES
	YES		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			
TRACEABLE TO NIST SRM 191I AND SRM 191II	YES		YES
	YES		
Sample ID			
Tested 03/05/24			
Tester BPHI			
Recorded 03/05/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

APPEARANCE AND COLOR	CLEAR TO PALE YELLOW LIQUID APPEARANCE	PASS
Sample ID		
Tested	03/05/24	
Tester	BPHI	
Recorded	03/05/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M103,T-24,GC,M106
Sample ID		
Tested	03/05/24	
Tester	BPHI	
Recorded	03/05/24	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	03/30/25
Sample ID		
Tested	03/05/24	
Tester	BPHI	
Recorded	03/05/24	
SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested	03/05/24	
Tester	BPHI	
Recorded	03/05/24	
Lot #	4022933	
Made	03/05/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager