

Page: 1

Date: 03/08/24 at 8:01 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 8628 SODIUM HYDROXIDE 0.1 NORMAL IN METHANOL					
Test	Target/UOM		Result		
Actual Normality	0.1000	Range 0.0995 - 0.1005	0.1000		
Actual Normanty	EQ/L	0.0000 0.1000	0.1000		
Sample ID					
Tested 03/08/	24				
Tester AFRE	24				
Recorded 03/08/	24				
QCSP7: STANDARDIZATION O			YES		
AQUEOUS BASES	METHOD		120		
	METHOD				
Sample ID	0.4				
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/			VE0		
N.I.S.T. Traceable to SRM 84 (K			YES		
	YES				
Sample ID					
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/					
PREPARED TO FORMULATION			YES		
TLE	YES				
Sample ID					
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/	24				
SAMPLE COLLECTION & HANG	DLING YES		YES		
PER LAB040	YES				
Sample ID					
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/	24				
DATE OF SAMPLING FOR TES			03/08/2024		
PURPOSES	MM/DD/YYYY		03/00/2024		
	171171/00/1111				
Sample ID	0.4				
Tested 03/08/	24				
Tester AFRE	0.4				
Recorded 03/08/		40.0 05.0	20.0		
AMBIENT TEMPERATURE DUF		10.0 - 35.0	20.0		
SAMPLING DEGC	DEG C				
Sample ID					
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/					
DATE OF TESTING OF SAMPLI	MM/DD/YYYY		03/08/2024		
	MM/DD/YYYY				
Sample ID					
Tested 03/08/	24				
Tester AFRE					
Recorded 03/08/	24				

03/08/24

Made

Page: 2

Date: 03/08/24 at 8:01 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

AMBIENT TEST TEMPE	RATURE	20.0 DEG C	15.0 - 30.0	20.0
Sample ID				
Tested	03/08/24			
Tester	AFRE			
Recorded	03/08/24			
APPEARANCE AND COI	LOR	CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	03/08/24			
Tester	AFRE			
Recorded	03/08/24			
INSTRUMENTS USED D	URING	INSTRUMENT		M103,GC,VF3-J7,BID-00488,PI
PREPARATION				D-M91
Sample ID				
Tested	03/08/24			
Tester	AFRE			
Recorded	03/08/24			
Actual Density		0.81669	0.81600 - 0.81727	0.81610
		GM/ML		
Sample ID				
Tested	03/08/24			
Tester	AFRE			
Recorded	03/08/24			
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	03/08/24			
Tester	AFRE			
Recorded	03/08/24			
Lot # 40306	19			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager