

Page: 1

Date: 03/15/24 at 1:49 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	6955			
		IUM CHLORIDE 0.1 MOLAR		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	03/15/24			
Tester	DTRA			
Recorded	03/15/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	03/15/24			
Tester	DTRA			
Recorded	03/15/24			
ACTUAL MOLARITY		0.1000	0.0991 - 0.1009	0.0996
		MOLARITY (M/L)		
Sample ID		- ( - ,		
Tested	03/15/24			
Tester	DTRA			
Recorded	03/15/24			
QCSP68:STANDARDIZ		YES		YES
SODIUM CHLORIDE N		QCSP68		
Sample ID		200.00		
Tested	03/15/24			
Tester	DTRA			
Recorded	03/15/24			
INSTRUMENTS USED I		INSTRUMENT		M55,M103,T56
PREPARATION		IN O THOUSEN		W60,W100,100
Sample ID Tested	03/15/24			
Tester	03/15/24 DTRA			
Recorded	03/15/24			
EXPIRATION DATE	00/10/24	MM/DD/YYYY		03/30/2025
EXPINATION DATE		MM/DD/YYYY		00/00/2020
Sample ID				
Tested	03/15/24			
Tester	DTRA			
Recorded	03/15/24			
Lot # 40314				
Made 03/15				
IVIAUE 00/1	J / L I			



Page: 2

Date: 03/15/24 at 1:49 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**