

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Prod	uct: ND121	F		
U		M CHLORIDE 1 NORMAL (1MOLA	AR)	
Test		Target/UOM	Range	Result
PREPARED TO FOR	RMULATION ON	YES	-	YES
FILE		YES		
Sample	e ID			
Tested	03/15/24			
Tester	JRIO			
Record	led 03/15/24			
CALCULATED MOLARITY		1.00	0.97 - 1.03	1.00
		MOL/L		
Sample	e ID			
Tested				
Tester	JRIO			
Record				
APPEARANCE AND COLOR		CLEAR TO SLIGHTLY TURBID		PASS
		LIQUID		
		APPEARANCE		
Sample	e ID			
Tested				
Tester	JRIO			
Record				
INSTRUMENTS USED DURING		INSTRUMENT		M87,T230
PREPARATION				
Sample	a ID			
Tested				
Tester	JRIO			
Record				
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Sample	םו ב			
Tested				
Tester	JRIO			
Record				
Lot # 40	31542			
	/15/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1