Page: 1

Date: 03/12/24 at 9:29 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	DC940F			
		STOCK STANDARD 100 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES	<u> </u>	YES
FILE		YES		
Sample ID				
Tested	03/12/24			
Tester	JENC			
Recorded	03/12/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE APPEARANCE		PASS
Sample ID				
Tested	03/12/24			
Tester	JENC			
Recorded	03/12/24			
INSTRUMENTS USED I	DURING	INSTRUMENT		GC,M115,VF1-Z35
PREPARATION				
Sample ID				
Tested	03/12/24			
Tester	JENC			
Recorded	03/12/24			
PPM CYANIDE 1 DEC PLACE		100.0	98.0 - 102.0	101.4
Sample ID				
Tested	03/12/24			
Tester	JENC			
Recorded	03/12/24			
QCSP47:STANDARDIZ	ATION	YES		YES
VARIOUS CYANIDE		QCSP47		
Sample ID				
Tested	03/12/24			
Tester	JENC			
Recorded	03/12/24			
Lot # 40312	284			
Made 03/12	2/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stol