



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 1

Date: 03/15/24 at 2:18 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 2182 COLOR STANDARD 5 UNITS (APHA)			
Test	Target/UOM	Range	Result
Prepared to formulation on File			
	Yes		YES
	YES		
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
Actual Color, APHA			
	5	4 - 6	5
	COLOR		
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
APPEARANCE AND COLOR			
	CLEAR PALE YELLOW		PASS
	APPEARANCE		
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		P173,P68,VF12-2
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
EXPIRATION DATE			
	MM/DD/YY		03/30/25
	MM/DD/YYYY		
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
SHELF LIFE OF PRODUCT			
	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested	03/15/24		
Tester	GWEB		
Recorded	03/15/24		
Lot # 4031404			
Made 03/15/24			



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 03/15/24 at 2:18 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager