03/12/24

Made

Page: 1

Date: 03/12/24 at 4:32 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Prod		M CHANDADD O O DDM EI/EI	TN TOOOGEANE	
	NITROGE.	N STANDARD 2.0 PPM W/W		<b>.</b> .
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Samp	le ID			
Teste	d 03/12/24			
Teste	r GWEB			
Recor	ded 03/12/24			
PPM NITROGEN V	V/W	2.00	1.90 - 2.10	2.00
		PPM		
Samp	le ID			
Teste				
Teste				
Recor				
INSTRUMENTS USED DURING		INSTRUMENT		P166,VF200-T18
PREPARATION				
Samp	le ID			
Teste				
Teste				
Recor				
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
	_	MM/DD/YYYY		03/33/2323
Samp	le ID			
Teste				
Teste				
Recor				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
				FAGG
6		APPEARANCE		
Samp				
Teste				
Teste				
Recor				
Lot # 40	031269			
0	2/10/04			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I