

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product				
	Chloride	e Std. 12,500 mg/L for	EZ4000/EZ4006 Series	
Test		Target/UOM	Range	Result
PREPARED TO FORM	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	03/13/24			
Tester	GWEB			
Recorded	03/13/24			
Calculated ppm w/v Chloride		12500	12495 - 12505	12500
		PPM W/V		
Sample ID				
Tested	03/13/24			
Tester	GWEB			
Recorded	03/13/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	03/13/24			
Tester	GWEB			
Recorded	03/13/24			
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF2-B11
PREPARATION				
Sample ID				
Tested	03/13/24			
Tester	GWEB			
Recorded	03/13/24			
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	03/13/24			
Tester	GWEB			
Recorded	03/13/24			
Lot # 4031	342			
Made 03/1	3/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA