Page: 1

Date: 03/13/24 at 11:42 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	UCT126F			
	CELITE	STD FOR TSS 40 MG/L SOL	UTION	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	03/13/24			
Tester	AROS			
Recorded	03/13/24			
ACTUAL PPM		40.00	39.50 - 40.50	40.08
		PPM		
Sample ID				
Tested	03/13/24			
Tester	AROS			
Recorded	03/13/24			
APPEARANCE AND COL	LOR	SLIGHTLY TURBID LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	03/13/24			
Tester	AROS			
Recorded	03/13/24			
INSTRUMENTS USED D		INSTRUMENT		M115,VF4-N22
PREPARATION				
Sample ID				
Tested	03/13/24			
Tester	AROS			
Recorded	03/13/24			
EXPIRATION DATE	00/10/24	MM/DD/YY		09/30/24
EXI INATION DATE		MM/DD/YYYY		00/00/24
Sample ID		WINNING DOTTER		
Tested	02/42/24			
Tester	03/13/24 AROS			
Recorded	03/13/24			
Lot # 4031112				
Made 03/13	/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I