Page: 1

Date: 03/22/24 at 11:08 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged P	roduct:	HON050F			
. aonagoa .		Sodium	Fluosilicate Standard So	olution	
Test			Target/UOM	Range	Result
PREPARED TO	FORMULA	TION ON	YES		YES
FILE			YES		
Sa	ample ID				
Te	ested 0	3/22/24			
Te	ester A	ROS			
Re	ecorded 0	3/22/24			
APPEARANCE AND COLOR			CLEAR WATER WHITE		PASS
			APPEARANCE		
Sa	ample ID				
Te	ested 0	3/22/24			
Te	ester A	ROS			
Re	ecorded 0	3/22/24			
INSTRUMENTS USED DURING		RING	INSTRUMENT		M115,VF1-Z31
PREPARATION	I				
Sa	ample ID				
Te	ested 0	3/22/24			
Te	ester A	ROS			
Re	ecorded 0	3/22/24			
EXPIRATION DATE			MM/DD/YY		09/30/24
			MM/DD/YYYY		
Sa	ample ID				
Te	ested 0	3/22/24			
Te	ester A	ROS			
Re	ecorded 0	3/22/24			
Lot #	4032209	9			
Made	03/22/2	24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager

hichard Stol