

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ				
_	BUFFER F	REFERENCE pH 5.00 ± 0.0		
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample				
Tested	03/19/24			
Tester	JRIO			
Recorde				
PH OF PRODUCT/SOLUTION		5.000	4.980 - 5.020	5.002
		PH		
Sample	ID			
Tested	03/19/24			
Tester	JRIO			
Recorde				
QCSP 20 PH CALIBR	RATION	YES		YES
METHODS		METHOD		
Sample	ID			
Tested	03/19/24			
Tester	JRIO			
Recorde				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample	ID			
Tested	03/19/24			
Tester	JRIO			
Recorde				
PH NIST TRACEABL		YES		YES
		YES		
Sample	חו			
Tested	03/19/24			
Tester	JRIO			
Recorde				
		YES		YES
NIST TRACEABLE TO SRM186I AND 186II		YES		125
	חו	123		
Sample				
Tested	03/19/24			
Tester	JRIO			
Recorde		VEC		VEO
PH NIST TRACEABL	E TO SRM 187	YES		YES
-		YES		
Sample				
Tested	03/19/24			
Tester	JRIO			
Recorde				
NSTRUMENTS USE	D DURING	INSTRUMENT		M103,GC,M108,T-28
PREPARATION				
Sample				
Tested	03/19/24			
Tester	JRIO			
Recorde	ed 03/19/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	03/30/25
Tester	03/19/24 JRIO 03/19/24	
Lot # 403120 Made 03/19,		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

hichard Stol

Quality Manager