03/20/24

Made

Page: 1

Date: 03/20/24 at 2:41 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

<b>Packaged Pro</b>	duct: AM346F			
	FLUORII	DE STANDARD 0.5 PPM IN 5	0% V/V TISAB	
Test		Target/UOM	Range	Result
PREPARED TO F	ORMULATION ON	YES		YES
FILE		YES		
Sam	ple ID			
Teste	ed 03/20/24			
Teste	er AROS			
Reco	orded 03/20/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Samı	ple ID			
Teste				
Teste	er AROS			
Reco	orded 03/20/24			
INSTRUMENTS U	SED DURING	INSTRUMENT		P78,VF1-Z16
PREPARATION				
Sam	ple ID			
Teste				
Teste	er AROS			
Reco	orded 03/20/24			
EXPIRATION DATE		MM/DD/YY		06/30/24
		MM/DD/YYYY		
Sam	ple ID			
Teste				
Teste	er AROS			
Reco				
SHELF LIFE OF PRODUCT		SHELF LIFE		3 MONTHS
		TIME		
Samı	ple ID			
Teste				
Teste				
Reco				
Lot # 4	032022			
_O( #				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St