Page: 1

Date: 03/22/24 at 3:17 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	ed Product	: ND672F			
			kygen Standard		
Test			Target/UOM	Range	Result
PREPARE	D TO FORMU	JLATION ON	YES	<u>.</u>	YES
FILE			YES		
	Sample ID				
	Tested	03/22/24			
	Tester	AROS			
	Recorded	03/22/24			
APPEARANCE AND COLOR		LOR	PALE YELLOW BROWN		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	03/22/24			
	Tester	AROS			
	Recorded	03/22/24			
INSTRUMENTS USED DURING		DURING	INSTRUMENT		M108,M100,VF1-Z25
PREPARA	ATION				
	Sample ID				
	Tested	03/22/24			
	Tester	AROS			
	Recorded	03/22/24			
EXPIRATION DATE			MM/DD/YYYY		07/30/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	03/22/24			
	Tester	AROS			
	Recorded	03/22/24			
Lo	ot # 40322	214			
Ma	nde 03/22	2/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager