Page: 1

Date: 03/28/24 at 3:03 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

_	-	Acetic Acid (KHP equivalent) MAKE FRESH		
Test	Target/UOM	Range	Result	
PREPARED TO FORMULATION O			YES	
FILE	YES			
Sample ID				
Tested 03/28/24				
Tester MMAT				
Recorded 03/28/24				
ACETIC ACID-KHP EQUIVALENT	15.0	14.5 - 15.5	15.1	
	PPM			
Sample ID				
Tested 03/28/24				
Tester MMAT				
Recorded 03/28/24				
APPEARANCE AND COLOR	CLEAR LIQUID		PASS	
	APPEARANCE			
Sample ID				
Tested 03/28/24				
Tester MMAT				
Recorded 03/28/24				
INSTRUMENTS USED DURING	INSTRUMENT		M108,VF1-Z28	
PREPARATION				
Sample ID				
Tested 03/28/24				
Tester MMAT				
Recorded 03/28/24				
EXPIRATION DATE	MM/DD/YYYY		06/30/2024	
-	MM/DD/YYYY			
Sample ID				
Tested 03/28/24				
Tester MMAT				
Recorded 03/28/24				
Lot # 4032822				
Made 03/28/24				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I