

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	8192			
		Standard 50 ppm 1ml=0.0)5mg	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES	~	YES
FILE		YES		
Sample ID				
Tested	03/25/24			
Tester	MMAT			
Recorded	03/25/24			
Calculated ppm silica		50.0	48.5 - 51.5	50.0
		PPM		
Sample ID				
Tested	03/25/24			
Tester	MMAT			
Recorded	03/25/24			
APPEARANCE AND COLOR		PASS		PASS
		APPEARANCE		
Sample ID				
Tested	03/25/24			
Tester	MMAT			
Recorded	03/25/24			
INSTRUMENTS USED	DURING	INSTRUMENT		P167,VF1-Z29
PREPARATION				
Sample ID				
Tested	03/25/24			
Tester	MMAT			
Recorded	03/25/24			
EXPIRATION DATE		MM/DD/YYYY		03/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	03/25/24			
Tester	MMAT			
Recorded	03/25/24			
Lot # 40322	221			
Made 03/2				
muuc				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA