

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produc	6494			
i ushuguu i ibuuu		EN STD 100 PPM $(1ML = 0.$	1MG N) SOLUTION	
Test		Target/UOM	Range	Result
PREPARED TO FORM		YES	nallye	YES
FILE		YES		125
Sample II	ר	120		
Tested	04/01/24			
Tester	AROS			
Recorded				
CERTIFIED PPM NITROGEN:		100.0	96.0 - 104.0	102.4
		PPM	00.0 104.0	102.4
Sample II	Ъ			
Tested	04/01/24			
Tester	AROS			
Recorded				
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample II	ר			
Tested	04/01/24			
Tester	AROS			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M108,GC,BOTTLES
PREPARATION				
Sample II	C			
Tested	04/01/24			
Tester	AROS			
Recorded				
EXPIRATION DATE		MM/DD/YYYY		07/30/2024
		MM/DD/YYYY		
Sample II	D			
Tested	04/01/24			
Tester	AROS			
Recorded	04/01/24			
Lot # 4032	2871			
	01/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1