Page: 1

Date: 04/02/24 at 1:23 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: SPE800F COD STANDARD 150 PPM  Test Target/UOM Range	Result
Test Target/UOM Range	
PREPARED TO FORMULATION ON YES	YES
<b>FILE</b> YES	
Sample ID	
Tested 04/02/24	
Tester MMAT	
Recorded 04/02/24	
<b>PPM COD</b> 150 137 - 162	149
PPM	
Sample ID	
Tested 04/02/24	
Tester MMAT	
Recorded 04/02/24	
APPEARANCE AND COLOR CLEAR LIQUID	PASS
APPEARANCE	
Sample ID	
Tested 04/02/24	
Tester MMAT	
Recorded 04/02/24	
	P124,P156,VF1-Z26
PREPARATION	1 121,1 100,11 1 220
Sample ID	
Tested 04/02/24	
Tester MMAT	
Recorded 04/02/24	
EXPIRATION DATE MM/DD/YYYY	07/30/2024
MM/DD/YYYY	07/30/2024
Sample ID	
Tested 04/02/24	
Tester MMAT	
Recorded 04/02/24	
Lot # 4040126	
Made 04/02/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I