



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: BAS717F			
	ANILINE HPLC MOBILE PHASE		
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
CALCULATED GPL	1.9823	1.9818 - 1.9828	1.9823
TETRABUTYLAMMONIUM HYDROG	GPL		
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
CALCULATED V/V% ACETONITRILE	41.0	40.8 - 41.3	41.0
	V/V%		
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
FILTERED THROUGH 0.2 UM FILTER	YES		YES
	YES		
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T-222,GC,M92
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
INITIAL PH OF SOLUTION	6.500	6.480 - 6.520	6.502
	PH		
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			
FINAL PH	6.500	6.300 - 6.900	6.698
	PH		
Sample ID			
Tested 04/04/24			
Tester JRIO			
Recorded 04/04/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

SHELF LIFE OF PRODUCT

6 MONTHS
 TIME

6 MONTHS

Sample ID

Tested 04/04/24

Tester JRIO

Recorded 04/04/24

EXPIRATION DATE

MM/DD/YYYY
 MM/DD/YYYY

10/30/2024

Sample ID

Tested 04/04/24

Tester JRIO

Recorded 04/04/24

Lot # 4040322

Made 04/04/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager