

Page: 1

Date: 04/02/24 at 1:27 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	SPX354F			
		NITROGE:	N STANDARD 10.0 PPM W/	J IN XYLENE	
Test			Target/UOM	Range	Result
PREPARED T	O FORMUL	ATION ON	YES		YES
FILE			YES		
5	Sample ID				
٦	Tested	04/02/24			
٦	Tester	AROS			
F	Recorded	04/02/24			
ASSAY OF PRODUCT, 2 DECIMAL		DECIMAL	10.00	9.95 - 10.05	10.00
PLACES			W/W%		
9	Sample ID				
	Tested	04/02/24			
1	Tester	AROS			
F	Recorded	04/02/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
5	Sample ID				
	rested	04/02/24			
-	Tester	AROS			
	Recorded	04/02/24			
INSTRUMENTS USED DURING			INSTRUMENT		P96,VF500-X44
PREPARATIO					
	Sample ID				
	rested	04/02/24			
	Tester	AROS			
	Recorded	04/02/24			
EXPIRATION			MM/DD/YY		04/30/25
			MM/DD/YYYY		
9	Sample ID				
	Tested	04/02/24			
	Tester	AROS			
	Recorded	04/02/24			
SHELF LIFE OF PRODUCT			1 YEAR		1 YEAR
V			TIME		
5	Sample ID				
	rested	04/02/24			
	Tester	AROS			
	Recorded	04/02/24			
Lot # 4040222					
Made 04/02/24					
····aac					



Page: 2

Date: 04/02/24 at 1:27 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**