

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Pr	roduct:	EQS522F			
<del>-</del>		AMMONIA	NITROGEN STANDARD 10.0	PPM AS N	
Test			Target/UOM	Range	Result
PREPARED TO	FORMULA	ATION ON	YES	<b>V</b>	YES
FILE			YES		
Sa	mple ID				
Tes	sted (	04/09/24			
Te	ster A	AROS			
Re	corded (	04/09/24			
CERTIFIED PPM NITRO		EN:	10.0	9.5 - 10.5	9.9
			PPM		
Sa	mple ID				
	-	04/09/24			
Tes	ster A	AROS			
Re	corded (	04/09/24			
APPEARANCE AND COLOR		DR	CLEAR LIQUID		PASS
			APPEARANCE		
Sa	mple ID				
	-	04/09/24			
Tes	ster A	AROS			
Re	corded (	04/09/24			
INSTRUMENTS USED DURING		RING	INSTRUMENT		P71,GC,VF2-B17
PREPARATION					
Sa	mple ID				
	-	04/09/24			
Tes		AROS			
Re	corded (	04/09/24			
EXPIRATION DATE			MM/DD/YYYY		05/14/2024
			MM/DD/YYYY		
Sa	mple ID				
		04/09/24			
Tes	ster A	AROS			
Re	corded (	04/09/24			
Lot #	404094	7			
Made	04/09/	24			
Maue	/				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA