

Page: 1

Date: 04/08/24 at 4:22 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## **Certificate of Analysis / QC Results**

Customer PO #:

Packaged Product:	1760			
	CERIC S	SULFATE TITRANT 0.1 NORM	$^{ m AL}$	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	<u> </u>	YES
FILE		YES		
Sample ID				
Tested	04/08/24			
Tester	SHAR			
Recorded	04/08/24			
Actual Normality		0.1000	0.0995 - 0.1005	0.0999
		EQ/L		
Sample ID				
Tested	04/08/24			
	SHAR			
Recorded	04/08/24			
Actual Density		1.08960	1.08631 - 1.09289	1.09019
-		GM/ML		
Sample ID				
	04/08/24			
	SHAR			
	04/08/24			
QCSP19:STANDARDIZAT	TION OF	YES		YES
CERIC SULFATE		QCSP19		
Sample ID				
	04/08/24			
	SHAR			
	04/08/24			
APPEARANCE AND COL		CLEAR ORANGE LIQUID		PASS
		APPEARANCE		
Sample ID				
	04/08/24			
	SHAR			
	04/08/24			
INSTRUMENTS USED DU		INSTRUMENT		M103,T-27,VF2-A14,M55
PREPARATION	=-			, , , , , ,
Sample ID				
	04/08/24			
	SHAR			
	04/08/24			
EXPIRATION DATE		MM/DD/YY		10/30/24
		MM/DD/YYYY		
Sample ID				
	04/08/24			
	SHAR			
	04/08/24			
Lot # 4040518				
_0,,,				
<b>Made</b> 04/08/	24			



Page: 2

Date: 04/08/24 at 4:22 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**