Page: 1

Date: 04/08/24 at 9:55 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

	TOTAL CY	YANIDE BLIND-QA STANDA	RD (ALL SAME LOT)	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	<u> </u>	YES
FILE		YES		
Sample ID				
Tested	04/08/24			
Tester	AROS			
Recorded	04/08/24			
Actual ppm Cyanide		0.40	0.09 - 0.51	0.40
		PPM		
Sample ID				
Tested	04/08/24			
Tester	AROS			
Recorded	04/08/24			
APPEARANCE AND COL	_OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	04/08/24			
Tester	AROS			
Recorded	04/08/24			
INSTRUMENTS USED D		INSTRUMENT		P145,GC,VF1-Z39
PREPARATION				2,2:2,
Sample ID				
Tested	04/08/24			
Tester	AROS			
Recorded	04/08/24			
EXPIRATION DATE	04/00/24	MM/DD/YYYY		04/22/2024
		MM/DD/YYYY		0-1/2L/2024
Sample ID		141141/ <i>DD</i> /1111		
Tested	04/08/24			
Tester	04/08/24 AROS			
	04/08/24			
Lot # 40408				
Made 04/08	/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland In