

Page: 1

Date: 04/09/24 at 3:03 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

## Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	EQS148F			
		C ACID 0.01 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	04/09/24			
Tester	MMAT			
Recorded	04/09/24			
Actual Normality		0.0100	0.0095 - 0.0105	0.0100
		EQ/L		
Sample ID				
Tested	04/09/24			
Tester	MMAT			
Recorded	04/09/24			
APPEARANCE AND COL	.OR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	04/09/24			
Tester	MMAT			
Recorded	04/09/24			
INSTRUMENTS USED D	URING	INSTRUMENT		P58,M100,M55
PREPARATION				
Sample ID				
Tested	04/09/24			
Tester	MMAT			
Recorded	04/09/24			
EXPIRATION DATE		MM/DD/YY		04/30/25
		MM/DD/YYYY		
Sample ID				
Tested	04/09/24			
Tester	MMAT			
	04/09/24			
SHELF LIFE OF PRODUC	СТ	SHELF LIFE		1 YEAR
		TIME		
Sample ID				
	04/09/24			
	MMAT			
Recorded	04/09/24	0.0000	0.00040 . 0.0000	0.000-0
Actual Density		0.99863	0.99843 - 0.99883	0.99872
o		GM/ML		
Sample ID	0.4/0.0/0.5			
Tested	04/09/24			
Tester	MMAT			
Recorded	04/09/24			
Lot # 4040951				
Made 04/09	/24			



Page: 2

Date: 04/09/24 at 3:03 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**