



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3540			
EDTA TITRANT 0.1 MOLAR (0.2N)			
Test	Target/UOM	Range	Result
Actual Molarity	0.1000 MOL/L	0.0995 - 0.1005	0.1000
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES QCSP4		YES
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
FORMULA WEIGHT	372.24 F.W.	= 372.24	372.24
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
Actual Density	1.01671 GM/ML	1.01605 - 1.01736	1.01654
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,T9,M55,M91,BID00515,TI D53329
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			
Traceable to N.I.S.T. SRM 915	YES YES		YES
Sample ID			
Tested 04/10/24			
Tester BPHI			
Recorded 04/10/24			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	04/30/26
Sample ID		
Tested	04/10/24	
Tester	BPHI	
Recorded	04/10/24	
SHELF LIFE OF PRODUCT	2 YEARS TIME	2 YEARS
Sample ID		
Tested	04/10/24	
Tester	BPHI	
Recorded	04/10/24	
Lot #	4040915	
Made	04/10/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager