

Page: 1

Date: 04/09/24 at 10:55 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product				
	Secondar	ry REF Std. pH 6.866 +	-0.010	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Sample ID				
Tested	04/09/24			
Tester	RSHE			
Recorded	04/09/24			
Calculated mol/L P0204A		0.0250	0.0247 - 0.0253	0.0250
		MOL/L		
Sample ID				
Tested	04/09/24			
Tester	RSHE			
Recorded	04/09/24			
Calculated mol/L P020		0.0250	0.0247 - 0.0253	0.0250
		MOL/L		
Sample ID				
Tested	04/09/24			
Tester	RSHE			
Recorded	04/09/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
7 27		APPEARANCE		
Sample ID		711 1 2711 0 11 0 2		
Tested	04/09/24			
Tester	RSHE			
Recorded	04/09/24			
		INSTRUMENT		M91,VF1-Z16
INSTRUMENTS USED DURING PREPARATION		INSTITUMENT		10191, 01 1-210
Sample ID Tested				
	04/09/24			
Tester	RSHE			
Recorded PATE	04/09/24	MM/DD/YYYY		05/09/2024
EXPIRATION DATE		MM/DD/YYYY		05/09/2024
Sample ID				
Tested	04/09/24			
Tester	RSHE			
Recorded	04/09/24			
Lot # 4040				
Made 04/0				
IVIAUE 9179	-, - ·			



Page: 2

Date: 04/09/24 at 10:55 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager