04/10/24

Made

Page: 1

Date: 04/10/24 at 1:57 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	6430			
		NITRATE	STD 1000 PPM N-NO3 (10	000 PPM NITROGEN FROM NITF	RATE)
Test			Target/UOM	Range	Result
PREPARED 1	TO FORMUL	ATION ON	YES		YES
FILE			YES		
:	Sample ID				
	Tested	04/10/24			
•	Tester	AROS			
	Recorded (04/10/24			
PPM NITROG	SEN FROM		1000.0	995.0 - 1005.0	1000.1
NITRATE(NO3)			PPM		
:	Sample ID				
	Tested (04/10/24			
	Tester	AROS			
	Recorded (04/10/24			
APPEARANCE AND COLOR		OR	CLEAR LIQUID		PASS
			APPEARANCE		
:	Sample ID				
	Tested (04/10/24			
•	Tester	AROS			
	Recorded (04/10/24			
INSTRUMENTS USED DURING		RING	INSTRUMENT		M108,VF1-Z31
PREPARATIO	ON				
:	Sample ID				
		04/10/24			
	Tester ,	AROS			
	Recorded (04/10/24			
EXPIRATION DATE			MM/DD/YYYY		08/30/2024
			MM/DD/YYYY		
	Sample ID				
		04/10/24			
	Tester	AROS			
		04/10/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St