04/18/24

Made

Page: 1

Date: 04/18/24 at 8:08 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	2213			
	COLOR S	TANDARD 75 HAZEN UNITS	(APHA, ASTM D1209)	
Test		Target/UOM	Range	Result
COLOR, APHA		75	73 - 77	74
		APHACOLOR		
Sample ID				
Tested	04/18/24			
Tester	AROS			
Recorded	04/18/24			
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	04/18/24			
Tester	AROS			
Recorded	04/18/24			
APPEARANCE AND COL	.OR	PALE YELLOW LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	04/18/24			
Tester	AROS			
Recorded	04/18/24			
INSTRUMENTS USED DU	JRING	INSTRUMENT		P174,VF2-A11,M82
PREPARATION				
Sample ID				
·	04/18/24			
Tester	AROS			
	04/18/24			
EXPIRATION DATE		MM/DD/YYYY		04/30/2025
		MM/DD/YYYY		
Sample ID				
•	04/18/24			
	AROS			
	04/18/24			
Lot # 404178	86			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St