Page: 1

Date: 04/15/24 at 9:11 AM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	ed Product:	AM132F		
	(	CHLORIDE STANDARD 10.0PPM	W/W	
Test		Target/UOM	Range	Result
PREPARE	D TO FORMULATION		<del>-</del>	YES
FILE		YES		
	Sample ID			
	Tested 04/1	15/24		
	Tester ARC	OS		
	Recorded 04/1	15/24		
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
	Sample ID			
	Tested 04/1	15/24		
	Tester AR0			
		15/24		
INSTRUM	ENTS USED DURIN	IG INSTRUMENT		P116,P46,VF5-L15
PREPARA	TION			
	Sample ID			
	·	15/24		
	Tester AR0			
		15/24		
PPM CHLORIDE		10.00	9.50 - 10.50	10.00
		PPM		
	Sample ID			
	·	15/24		
	Tester ARC			
		15/24		
EXPIRATION DATE		MM/DD/YY		04/30/25
		MM/DD/YYYY		
	Sample ID	· · · · ·		
	·	15/24		
	Tester ARC			
Ì		15/24		
	ot # 4041141			
Ma	ide 04/15/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.