

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

| Test | Target/UOM | Range | Result |
|------------------------|---------------|----------------|----------|
| PREPARED TO FORMULATI | | | YES |
| FILE | YES | | |
| Sample ID | | | |
| Tested 04/ | 15/24 | | |
| Tester BPI | -11 | | |
| Recorded 04/ | 15/24 | | |
| CALCULATED MOLARITY | 1.000 | 0.992 - 1.008 | 1.000 |
| | MOL/L | | |
| Sample ID | | | |
| Tested 04/ | 15/24 | | |
| Tester BPI | | | |
| | 15/24 | | |
| APPEARANCE AND COLOR | PALE YELLOW | | PASS |
| | APPEARANCE | | |
| Sample ID | | | |
| | 15/24 | | |
| Tester BPI | | | |
| | 15/24 | | |
| INSTRUMENTS USED DURIN | | | M87,T165 |
| PREPARATION | | | |
| Sample ID | | | |
| • | 15/24 | | |
| Tester BPI | | | |
| | 15/24 | | |
| PH OF PRODUCT/SOLUTION | 9 .900 | 9.800 - 10.000 | 9.835 |
| | PH | | |
| Sample ID | | | |
| | 15/24 | | |
| Tester BPI | | | |
| | 15/24 | | |
| EXPIRATION DATE | MM/DD/YY | | 10/30/24 |
| | MM/DD/YYYY | | |
| Sample ID | | | |
| | 15/24 | | |
| Tester BPI | | | |
| | 15/24 | | |
| SHELF LIFE OF PRODUCT | 6 MONTHS | | 6 MONTHS |
| | TIME | | |
| Sample ID | | | |
| Tested 04/ | 15/24 | | |
| Tester BPI | 41 | | |
| Recorded 04/ | 15/24 | | |



Page: 2 Date: 04/15/24 at 11:35 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager