

Page: 1

Date: 04/17/24 at 3:18 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: CY076F			
-		IUM PERMANGANATE 0.02 NO	'MAL	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	04/17/24			
Tester	JENC			
Recorded	04/17/24			
Actual Normality		0.0200 EQ/L	0.0196 - 0.0204	0.0201
Camania ID		EQ/E		
Sample ID				
Tested	04/17/24			
Tester	JENC 04/47/24			
Recorded	04/17/24	CLEAR RURRIE LIQUIR		DACC
APPEARANCE AND CO	DLOR	CLEAR PURPLE LIQUID APPEARANCE		PASS
Sample ID				
Tested	04/17/24			
Tester	JENC			
Recorded	04/17/24			
INSTRUMENTS USED I	DURING	INSTRUMENT		VF200-T13,VF1-Z34
PREPARATION				
Sample ID				
Tested	04/17/24			
Tester	JENC			
Recorded	04/17/24			
Actual Density		0.99868 GM/ML	0.99857 - 0.99880	0.99870
0 1 10		GW/WL		
Sample ID				
Tested	04/17/24			
Tester	JENC			
Recorded	04/17/24	MANDENA		04/00/05
EXPIRATION DATE		MM/DD/YY MM/DD/YYYY		04/30/25
Sample ID				
Tested	04/17/24			
Tester	JENC			
Recorded	04/17/24			
Lot # 4041	729			
Made 04/1	7/24			

Page: 2

Date: 04/17/24 at 3:18 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager