

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6938			
. donagoa i roadon		TIVITY STANDARD 5.0 uMHOS		
Test		Target/UOM	Range	Result
PREPARED TO FORMU		YES	Nange	YES
FILE		YES		
Sample ID				
Tested	04/16/24			
Tester	BPHI			
Recorded	04/16/24			
CONDUCTIVITY		5.0	4.8 - 5.2	5.0
		UMHOS		
Sample ID				
Tested	04/16/24			
Tester	BPHI			
Recorded	04/16/24			
QCSP30:CONDUCTIVIT		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	04/16/24			
Tester	BPHI			
Recorded	04/16/24			
TRACEABLE TO N.I.S.T.		YES		YES
		YES		
Sample ID				
Tested	04/16/24			
Tester	BPHI			
Recorded	04/16/24			
APPEARANCE AND COI		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	04/16/24			
Tester	BPHI			
Recorded	04/16/24			
INSTRUMENTS USED D		INSTRUMENT		M92,T234
PREPARATION	orano			1102,1201
Sample ID				
Tested	04/16/24			
Tester	04/16/24 BPHI			
Recorded	04/16/24			
EXPIRATION DATE	07/10/24	MM/DD/YY		07/30/24
		MM/DD/YYYY		
Sample ID				
Tested	04/16/24			
Tester	04/16/24 BPHI			
Recorded	04/16/24			
Made 04/16	/ 2 4			



Page: 2 Date: 04/16/24 at 10:20 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager