

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged		148F		
	SUL	FURIC ACID 0.01 NORMAL		
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON				YES
FILE		YES		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
Actual Norm	nality	0.0100	0.0095 - 0.0105	0.0100
		EQ/L		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
APPEARAN	CE AND COLOR	CLEAR LIQUID		PASS
		APPEARANCE		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
INSTRUMEN	ITS USED DURING	INSTRUMENT		P123,M100
PREPARAT	ON			
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
EXPIRATIO	N DATE	MM/DD/YY		04/30/25
		MM/DD/YYYY		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
SHELF LIFE OF PRODUCT		SHELF LIFE		1 YEAR
	-	TIME		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
Actual Dens		0.99863	0.99843 - 0.99883	0.99860
		GM/ML		
	Sample ID			
	Tested 04/19/24			
	Tester AROS			
	Recorded 04/19/24			
Lot				
Mad	e 04/19/24			



Page: 2 Date: 04/19/24 at 6:05 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

ichard Stop

Quality Manager