

Page: 1

Date: 05/01/24 at 11:01 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 693	38		
	NDUCTIVITY STANDARD 5.0 uMHOS		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION		- 3 -	YES
FILE	YES		
Sample ID			
Tested 05/01/2	24		
Tester JRIO			
Recorded 05/01/2	24		
CONDUCTIVITY	5.0	4.8 - 5.2	5.0
	UMHOS		
Sample ID			
Tested 05/01/2	24		
Tester JRIO			
Recorded 05/01/2	24		
QCSP30:CONDUCTIVITY	YES		YES
STD&MEASUREMENT	QCSP30		
Sample ID			
Tested 05/01/2	24		
Tester JRIO			
Recorded 05/01/2	24		
TRACEABLE TO N.I.S.T. SRM 3	191 YES		YES
	YES		
Sample ID			
Tested 05/01/2	24		
Tester JRIO			
Recorded 05/01/2	24		
APPEARANCE AND COLOR	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 05/01/2	24		
Tester JRIO			
Recorded 05/01/2	24		
INSTRUMENTS USED DURING	INSTRUMENT		M-92,T-234
PREPARATION			0=,1 =0 .
Sample ID			
Tested 05/01/2	24		
Tester JRIO	. 1		
Recorded 05/01/2	24		
EXPIRATION DATE	MM/DD/YY		08/30/24
EXI IIIAIION DATE	MM/DD/YYYY		00/00/24
Sample ID			
Tested 05/01/2	24		
Tester JRIO	-T		
Recorded 05/01/2	24		
Lot # 4050124			
Made 05/01/24			



Page: 2

Date: 05/01/24 at 11:01 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager