04/30/24

Made

Page: 1

Date: 04/30/24 at 3:02 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged F	Product:	1792			
		CHLORIN	E STANDARD 100 PPM AS	CL2 CERTIFIED SOLUTION	
Test			Target/UOM	Range	Result
Actual Density	у		0.99951	0.99884 - 1.00018	0.99936
_			GM/ML		
	Sample ID				
		04/30/24			
T		AFRE			
R	Recorded (	04/30/24			
INSTRUMENTS USED DURING		INSTRUMENT		VF12-1,M55,BID00488,PIDP16	
PREPARATIO	N				7
S	Sample ID				
T	ested (	04/30/24			
T	ester /	AFRE			
R	Recorded (	04/30/24			
EXPIRATION DATE		MM/DD/YY		07/05/24	
			MM/DD/YYYY		
s	Sample ID				
	-	04/30/24			
		AFRE			
		04/30/24			
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS	
			TIME		
s	Sample ID				
		04/30/24			
		AFRE			
		04/30/24			
PPM CHLORIDE		3 1/00/21	100.00	98.00 - 102.00	100.17
011201112			PPM	00.00 102.00	100.17
s	Sample ID				
	•	04/30/24			
		4/30/24 4FRE			
		04/30/24			
	404293				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland St