

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Package	d Product:				
		COD STA	NDARD 500 PPM W/W AQUEC	US	
Test			Target/UOM	Range	Result
PREPAREI	D TO FORMU	LATION ON	YES	-	YES
FILE			YES		
	Sample ID				
	Tested	04/29/24			
	Tester	DTRA			
	Recorded	04/29/24			
PPM COD			500	485 - 515	489
			PPM		
	Sample ID				
	Tested	04/29/24			
	Tester	DTRA			
	Recorded	04/29/24			
APPEARANCE AND COLOR		LOR	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	04/29/24			
	Tester	DTRA			
	Recorded	04/29/24			
INSTRUMENTS USED DURING		OURING	INSTRUMENT		M108,VF2-B37,M109
PREPARA	ΓΙΟΝ				
	Sample ID				
	Tested	04/29/24			
	Tester	DTRA			
	Recorded	04/29/24			
EXPIRATION DATE			MM/DD/YYYY		07/04/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	04/29/24			
	Tester	DTRA			
	Recorded	04/29/24			
Lo	t# 40426	548			
Ма					
inat					

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA