

Page: 1

Date: 04/30/24 at 9:40 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	6950			
CONDUCTIVITY STANDARD 1,000 uMHOS				
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES	<u> </u>	YES
FILE		YES		
Sample ID				
	04/30/24			
Tester	BPHI			
	04/30/24			
CONDUCTIVITY @25 C.	04/30/24	1000	999 - 1001	1000
		US/CM	999 - 1001	1000
Sample ID				
Tested	04/30/24			
Tester	BPHI			
Recorded	04/30/24			
CSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		-
Sample ID		400.00		
	04/00/04			
	04/30/24			
Tester	BPHI			
	04/30/24			
TRACEABLE TO N.I.S.T.	SRM 3191	YES		YES
		YES		
Sample ID				
·	04/30/24			
Tester	BPHI			
	04/30/24			
APPEARANCE AND COL		CLEAR WATER WHITE		PASS
MEFEMBAINGE AND COL	.un			FAGG
		LIQUID		
		APPEARANCE		
Sample ID				
Tested	04/30/24			
Tester	BPHI			
	04/30/24			
INSTRUMENTS USED DI		INSTRUMENT		M-92,T-229
PREPARATION				0=,. ==0
Sample ID	0.4/0.5/5			
Tested	04/30/24			
Tester	BPHI			
	04/30/24			
EXPIRATION DATE		MM/DD/YY		04/30/25
		MM/DD/YYYY		
Sample ID				
•	04/30/24			
Tester	BPHI			
	04/30/24	CHELETIEE		4 \/E 4 D
SHELF LIFE OF PRODU	6 1	SHELF LIFE TIME		1 YEAR
Sample ID				
	04/30/24			
Tester	BPHI			
Recorded	04/30/24			

Page: 2

Date: 04/30/24 at 9:40 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 4043019 Made 04/30/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager