

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Proc	luct: 2161			
COD STANDARD 100 PPM SOLUTION				
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES		YES
FILE		YES		
Samp	le ID			
Tested	d 04/29/24			
Tester	r DTRA			
Recor	ded 04/29/24			
PPM COD		100.0	97.0 - 103.0	99.3
		PPM		
Samp	le ID			
Tested				
Tester				
Recor				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Samp	ID ID			
Tested				
Tester				
Recor				
INSTRUMENTS USED DURING		INSTRUMENT		P126,VF1-Z42,M109
PREPARATION	DURING	INSTROMENT		1 120, 11 1-242, 10103
Samp				
Testeo	• = • . = .			
Tester				
Recor				
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS
		TIME		
Samp				
Testeo	• = • . = .			
Tester				
Recor				
EXPIRATION DATE		MM/DD/YYYY		07/04/2024
		MM/DD/YYYY		
Samp				
Testeo	d 04/29/24			
Tester				
Recor	ded 04/29/24			
Lot # 40)42610			
	1/29/24			
Made 04	1/23/24			



Page: 2 Date: 04/29/24 at 2:35 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard Stop

Quality Manager