



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: S2935			
SODIUM HYDROXIDE TECHNICAL FLAKE			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 96.6%		98.684%
	%		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
IRON	MAX 20.0 PPM		4.0868 PPM
	PPM		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
SODIUM CARBONATE	MAX 0.3 - 0.4%		0.2513%
	%		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
SODIUM CHLORIDE	MAX 0.03 - 0.04%		0.005263%
	%		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
EXPIRATION DATE	MM/DD/YY		11/22/24
	MM/DD/YYYY		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
SHELF LIFE OF PRODUCT	SHELF LIFE		2 YEARS
	TIME		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
MANUFACTURE DATE	MM/DD/YY		12/02/22
	MM/DD/YYYY		
Sample ID			
Tested	06/26/23		
Tester	TBUR		
Recorded	06/26/23		
Lot #	317401		
Made	06/23/23		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 06/26/23 at 9:24 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager