



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8375			
SILVER NITRATE 0.01 NORMAL IN IPA (ASTM D3227-16)			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID	YES		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		05/15/2024
Sample ID	MM/DD/YYYY		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	21.0
Sample ID	DEG C		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		05/15/2024
Sample ID	MM/DD/YYYY		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0	15.0 - 30.0	21.0
Sample ID	DEG C		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>Actual Normality</b>	0.0100	0.0097 - 0.0103	0.0102
Sample ID	EQ/L		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			
<b>Actual Density</b>	0.79067	0.79000 - 0.79135	0.79022
Sample ID	GM/ML		
Tested 05/15/24			
Tester DTRA			
Recorded 05/15/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>QCSP54:STANDARDIZATION OF LOW NORMAL SN</b>	YES QCSP54	YES
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>N.I.S.T. Traceable to SRM 919 NaCl</b>	YES YES	YES
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>IPA TESTED FOR PEROXIDES</b>	PASSES YES	PASSES
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE	PASS
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M103,EF4-JH,T-211,M55,BID5 2
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	05/30/25
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		
<b>SHELF LIFE OF PRODUCT</b>	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested 05/15/24		
Tester DTRA		
Recorded 05/15/24		

**Lot #** 4051454

**Made** 05/15/24



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 05/15/24 at 5:03 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**